

Instrument List Three Gas Recirculation

(Flammable and not flammable Mixer)

Numbering	Description	Supplier	Model Supplier Nb.	Range	Output Signal	Supply Voltage
PT-1103 1203 1303	Pressure transmitter	Sensortechnics	CTE8005GQ4	0-5 bar	4-20 mA	24VDC
		TRAFAG AG	885276P515041958 Ex Series	0-4 bar	4-20 mA	24VDC
PT-1009	Pressure transmitter	Sensortechnics	CTE8002GQ4	0-2 bar	4-20 mA	24VDC
		TRAFAG AG	885275P515041958 Ex Series	0-2.5 bar	4-20 mA	24VDC
YV-1004 1008 1105 1108 1205 1208 1305 1308	Solenoid valve	Lucifer	321 K 41 06 BE			24VDC
			Ex Series			24VDC
			321 K 45 06 BE			24VDC
			Ex Series			24VDC
YV-1006	Solenoid valve	Lucifer	321 K 46 06 BE			24VDC
			Ex Series			24VDC
XMFC-1106	Mass flow controler	BRONKHORST	DMFC Serials 201C, D 201AC 202AC, D 203AC		Profibus DP	24VDC
XMFC-1109	Mass flow controler	BRONKHORST	DMFC Serials 201C, D 201AC 202AC, D 203AC		Profibus DP	24VDC
XMFC-1206	Mass flow controler	BRONKHORST	DMFC Serials 201C, D 201AC 202AC, D 203AC		Profibus DP	24VDC
XMFC-1209	Mass flow controler	BRONKHORST	DMFC Serials 201C, D 201AC 202AC, D 203AC		Profibus DP	24VDC
XMFC-1306	Mass flow controler	BRONKHORST	DMFC Serials 201C, D 201AC 202AC, D 203AC		Profibus DP	24VDC
XMFC-1309	Mass flow controler	BRONKHORST	DMFC Serials 201C, D 201AC 202AC, D 203AC		Profibus DP	24VDC

Instrument List Three Gas Recirculation

(Flammable and not flammable Mixer)

ATEX / Ex.
Non
Yes
Non
Yes
Non
Yes
Non
Yes
Non
Yes
Non
Non
Non
Non
Non
Non